



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
[www.uspto.gov](http://www.uspto.gov)

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/635,319	08/06/2003	Benjamin Garden	MSDI-67/PC933.00	8113
52196	7590	09/12/2007	EXAMINER	
KRIEG DEVAULT LLP			SWIGER III, JAMES L	
ONE INDIANA SQUARE, SUITE 2800			ART UNIT	PAPER NUMBER
INDIANAPOLIS, IN 46204-2709			3733	
MAIL DATE		DELIVERY MODE		
09/12/2007		PAPER		

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

<b>Office Action Summary</b>	<b>Application No.</b>	<b>Applicant(s)</b>
	10/635,319	GARDEN ET AL.
	<b>Examiner</b>	<b>Art Unit</b>
	James L. Swiger	3733

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

#### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

#### Status

- 1) Responsive to communication(s) filed on 11 June 2007.
- 2a) This action is FINAL.                    2b) This action is non-final.
- 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

#### Disposition of Claims

- 4) Claim(s) 1-31 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) Claim(s) \_\_\_\_\_ is/are allowed.
- 6) Claim(s) 1-31 is/are rejected.
- 7) Claim(s) \_\_\_\_\_ is/are objected to.
- 8) Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

#### Application Papers

- 9) The specification is objected to by the Examiner.
- 10) The drawing(s) filed on 8/6/2003 is/are: a) accepted or b) objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

#### Priority under 35 U.S.C. § 119

- 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) All    b) Some \* c) None of:
  1. Certified copies of the priority documents have been received.
  2. Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

#### Attachment(s)

- |  |   |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892)                     | 4) <input type="checkbox"/> Interview Summary (PTO-413)           |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date: _____                                      |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)          | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date: _____   | 6) <input type="checkbox"/> Other: _____                          |

## DETAILED ACTION

### ***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims s 1-16 and 18-31 are rejected under 35 U.S.C. 102(b) as being anticipated by Michelson (US Patent 6,193,721). Michelson discloses a spinal stabilization system having a stabilization device that is a plate (see Fig. 1) positionable along the spinal column, and an auxiliary element (60) associated herewith and capable of moving relative to the plate, and a holding element with a proximal (92), distal (56) and intermediate portions (90/62) that is cannulated and unthreaded along the longitudinal axis through the auxiliary element. It may also be considered to have a 'distally oriented' engagement surface. The proximal and distal elements are also in cooperation with the auxiliary portion maintain a position along the spine. The plate also has bone anchor openings (6), and wherein the auxiliary device has a retaining element that at least partially overlaps the openings (see radially outward notched portion between 56 and 94 in Fig. 33, approx 71). The retaining device also has a fastener (66) that fits into the cannulation with a central aperture (Fig. 32) that also aligns with the aperture of the plate (see Fig. 34). Michelson also discloses at least a first and second driving tool portions (see Fig. 38) which also have a recess (142).

***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claim 17 is rejected under 35 U.S.C. 103(a) as being unpatentable over Michelson '721 in view of Kuslich et al. (US Publication 2003/0083749). Michelson discloses the claimed invention except for the teaching of the use of the device as a corpectomy implant. Kuslich et al. teaches the use of a stabilization system that is, more specifically, a corpectomy device (pars. 0030 through 0035). It would have been obvious to one having ordinary skill in the art at the time the invention was made to construct the device of Michelson that may also be used as a Corpectomy device in view of Kuslich et al. to better use the device to stabilize the spine.

***Response to Arguments***

Applicant's arguments filed 6/11/2007 have been fully considered but they are not persuasive. It is still held that the prior art of reference reads on the claims. Regarding the claim amendments, the auxiliary device may be considered "mounted thereto" when the system is configured for insert. The distal end of the device extends through a cannulation in the auxiliary device when it is used with the stabilization device. Further, the proximal portion of the holding element, attached, is capable of

being "operable to rotate" the auxiliary element in a manner at least considered relative to the said stabilization device at a distal portion.

***Conclusion***

**THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to James L. Swiger whose telephone number is 571-272-5557. The examiner can normally be reached on Monday through Friday, 9:00am to 5:30pm.


If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on 571-272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3733

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

 9/3/07

JLS



EDUARDO ROBERT  
SUPERVISORY PATENT EXAMINER